Australian Standard™

Safety in laboratories

Part 4: Ionizing radiations

This Australian Standard was prepared by Committee, CH/26—Safety in Laboratories. It was approved on behalf of the Council of Standards Australia on 1 May 1998 and published on 5 July 1998.

The following interests are represented on Committee CH/26: Agriculture Victoria Australian Council of Trade Unions CSIRO Manufacturing Science and Technology Department of Defence, Australia Department of Labour, New Zealand Environmental Science and Research, New Zealand ESR Kenepuru Science Centre, New Zealand Ministry of Commerce, New Zealand National Association of Testing Authorities, Australia New Zealand Chemical Industry Council New Zealand Microbiological Society Victorian WorkCover Authority WorkCover New South Wales

This Standard was prepared by the following participants:

Environment Protection Authority of N.S.W.

Independent consultant health physicists

University of New South Wales

University of Western Australia

This Standard was issued in draft form for comment as DR 96063.

Review of Australian Standards. To keep abreast of progress in industry, Australian Standards are subject to periodic review and are kept up to date by the issue of amendments or new editions as necessary. It is important therefore that Standards users ensure that they are in possession of the latest edition, and any amendments thereto.

Full details of all Australian Standards and related publications will be found in the Standards Australia Catalogue of Publications; this information is supplemented each month by the magazine 'The Australian Standard', which subscribing members receive, and which gives details of new publications, new editions and amendments, and of withdrawn Standards.

Suggestions for improvements to Australian Standards, addressed to the head office of Standards Australia, are welcomed. Notification of any inaccuracy or ambiguity found in an Australian Standard should be made without delay in order that the matter may be investigated and appropriate action taken.

Australian Standard™

Safety in laboratories

Part 4: Ionizing radiations

Originated as AS 2243.4—1979. Previous edition AS 2243.4(Int)—1994. Revised and designated AS 2243.4—1998.

Published by Standards Australia (Standards Association of Australia) 1 The Crescent, Homebush, NSW 2140

This is a free 8 page sample. Access the full version online.

2

This Standard was prepared by Subcommittee CH/26/4, Ionizing Radiation, on behalf of the Joint Standards Australia/Standards New Zealand Committee CH/26, Safety in Laboratories, to supersede AS 2243.4(Int)—1994.

This Standard is the result of a consensus among representatives on the Joint Committee to produce it as an Australian Standard.

The objective of this Standard is to promote safe working practices and to prevent unnecessary exposure of persons working in laboratories containing sources of ionizing radiation.

This Standard was prepared after consideration of the *Recommendations for limiting exposure to ionizing radiation* (1995) [NOHSC:3022] and *National standard for limiting occupational exposure to ionizing radiation* [NOHSC:1013 (1995)] (Radiation Health Series No. 39) issued by the National Health and Medical Research Council (NHMRC) and Worksafe Australia, and the most recent recommendations of the International Commission on Radiological Protection (ICRP). Where necessary, future relevant publications of the NHMRC and ICRP should be consulted as they become available.

The sections of this Standard dealing with dose limits, accident and emergency procedures and the requirements for radiological laboratories have been further refined since the superseded interim edition. Where there is a choice of data, the most restrictive data is presented. For more detailed information, the referenced source documents should be consulted. The previous edition contained some design information for laboratories which has been omitted from this Standard and may now be found in AS/NZS 2982.1, *Laboratory design and construction*, Part 1: *General requirements*.

This Standard is Part 4 in a series aimed at promoting safety in laboratories. The series is as follows:

- Part 1: General
- Part 2: Chemical aspects
- Part 3: Microbiology
- Part 4: Ionizing radiations (this Standard)
- Part 5: Non-ionizing radiations
- Part 6: Mechanical aspects
- Part 7: Electrical aspects
- Part 8: Fume cupboards
- Part 9: Recirculating fume cabinets
- Part 10: Storage of chemicals

Statements expressed in mandatory terms in notes to tables are deemed to be requirements of this Standard.

The terms 'normative' and 'informative' have been used in this Standard to define the application of the appendix to which they apply. A 'normative' appendix is an integral part of a Standard, whereas an 'informative' appendix is only for information and guidance.

CONTENTS

	F	` age
FOREW	'ORD	5
SECTIC	N 1 SCOPE AND GENERAL	
11	SCOPE	6
1.1	REFERENCED DOCUMENTS	6
1.3	DEFINITIONS	6
SECTIC	N 2 HAZARDS AND THEIR CONTROL	
2.1	HAZARDS FROM IONIZING RADIATION AND THEIR CONTROL	11
2.2	LEGISLATION	11
2.3	TRAINING	11
SECTIC	N 3 RESPONSIBILITIES	
3.1	GENERAL	12
3.2	RESPONSIBILITIES OF THE MANAGEMENT/EMPLOYER	12
3.3	RADIATION PROTECTION ADVISER (RPA)	12
3.4	OCCUPATIONALLY EXPOSED PERSONS	13
SECTIC	ON 4 PRIMARY DOSE LIMITS	
4.1	GENERAL	15
4.2	NHMRC RECOMMENDATIONS	15
SECTIC	ON 5 CLASSIFICATION AND MONITORING OF PERSONS	
	EXPOSED TO IONIZING RADIATIONS	
5.1	OCCUPATIONALLY EXPOSED PERSONS	19
5.2	MEMBERS OF THE PUBLIC	19
SECTIC	N 6 IONIZING RADIATION PROTECTION	
6.1	GENERAL	20
6.2	DESIGNATED RADIATION AREAS (DRAS)	21
6.3	RADIATION WARNING SIGNS	21
6.4	SHIELDING	21
6.5	PERSONS WORKING IN ISOLATION	21
6.6	IONIZING RADIATION PROTECTION—SEALED SOURCES	
	AND RADIATION-PRODUCING APPARATUS	22
6.7	IONIZING RADIATION PROTECTION—UNSEALED	
	RADIOACTIVE SUBSTANCES	25
SECTIC	N 7 STORAGE AND TRANSPORT	
7.1	STORAGE OF RADIOACTIVE SUBSTANCES	29
7.2	TRANSPORT OF RADIOACTIVE SUBSTANCES	30

SECTI	ON 8 DISPOSAL OF RADIOACTIVE WASTES	
8.1	GENERAL	31
8.2	SOLID WASTE	31
8.3	LIQUID WASTE	31
8.4	AIRBORNE WASTE	33
8.5	OPTIONS FOR THE DISPOSAL OF UNWANTED SEALED SOURCES	33
SECTI	ON 9 MONITORING IONIZING RADIATION	
9.1	GENERAL	35
9.2	CALIBRATION OF MONITORING INSTRUMENTS	35
9.3	PERSONAL MONITORING	35
9.4	RADIATION DOSE RECORD	36
9.5	AREA MONITORING	36
SECTI	ON 10 TRAINING, INSTRUCTION, ACCIDENTS AND	
	EMERGENCY PROCEDURES	
10.1	TRAINING AND INSTRUCTION	37
10.2	ACCIDENTS AND EMERGENCY PROCEDURES	37
SECTI	ON 11 DESIGN REQUIREMENTS FOR RADIATION,	
	RADIOISOTOPE AND RADIOLOGICAL LABORATORIES	
11.1	GENERAL	41
11.2	PLANNING BRIEF	41
11.3	TYPES OF HAZARDS	42
11.4	DESIGN AND CONSTRUCTION REQUIREMENTS	42
11.5	SUMMARY OF DESIGN REQUIREMENTS	42
APPEN	NDICES	
Α	CONTROL OF RADIATION EXPOSURE OF THE PUBLIC	44
В	REFERENCED AND RELATED DOCUMENTS	51
С	DERIVED RADIATION PROTECTION STANDARDS	55
D	MEASUREMENT OF IONIZING RADIATIONS	66
Е	IONIZING RADIATION SYMBOL	70
F	RADIOISOTOPE LABORATORY GRADING	71
G	MANAGEMENT OF SOLID, LIQUID AND GASEOUS RADIOACTIVE	
	WASTE	72
Н	TRANSPORT OF RADIOACTIVE SUBSTANCES	77

© Copyright – STANDARDS AUSTRALIA

Users of Standards are reminded that copyright subsists in all Standards Australia publications and software. Except where the Copyright Act allows and except where provided for below no publications or software produced by Standards Australia may be reproduced, stored in a retrieval system in any form or transmitted by any means without prior permission in writing from Standards Australia. Permission may be conditional on an appropriate royalty payment. Requests for permission and information on commercial software royalties should be directed to the head office of Standards Australia.

Standards Australia will permit up to 10 percent of the technical content pages of a Standard to be copied for use exclusively in-house by purchasers of the Standard without payment of a royalty or advice to Standards Australia.

Standards Australia will also permit the inclusion of its copyright material in computer software programs for no royalty payment provided such programs are used exclusively in-house by the creators of the programs.

Care should be taken to ensure that material used is from the current edition of the Standard and that it is updated whenever the Standard is amended or revised. The number and date of the Standard should therefore be clearly identified.

The use of material in print form or in computer software programs to be used commercially, with or without payment, or in commercial contracts is subject to the payment of a royalty. This policy may be varied by Standards Australia at any time.

FOREWORD

Radiation is the emission of energy, in the form of rays, wave motion or particles, from a source. The various kinds of radiation can be subdivided broadly as follows:

- (a) *Ionizing radiations*—radiations that by reason of their nature and energy, interact with matter causing ionization by removing or adding electrons from or to the atoms of substances through which the radiations pass, including those that comprise the human organism. Sufficiently energetic radiation may also cause changes in the nuclei of the atoms in the substance. Ionizing radiation comprises X-rays, gamma rays, alpha particles, beta particles (i.e. high energy electrons), neutrons, protons and other nuclear particles. Ultraviolet radiation of wavelength less than 100 nm will also cause ionization in air.
- (b) *Non-ionizing radiations*—radiations that do not cause ionization or nuclear changes, but can harm the human body in other ways. Potentially harmful non-ionizing radiations are propagated as electromagnetic waves (in particular, when the wavelength is between about 100 nm and 1 km), and acoustic noise (i.e. unwanted sound).

This Standard does not deal with non-ionizing radiation, which is covered in AS 2243.5, *Safety in laboratories*, Part 5: *Non-ionizing radiations*.

NOTES:

- 1 Although referred to in Item (a), ultraviolet radiation is generally classified as non-ionizing radiation and is therefore considered in AS 2243.5.
- 2 Radiation protection is concerned with the protection of individuals, their progeny, the human race as a whole and the environment, while still allowing necessary activities from which radiation exposure might result.
- 3 It should be noted that various forms of high voltage apparatus may produce ionizing radiations. Various chemicals and minerals may be radioactive by virtue of the fact that they contain traces of naturally occurring radioactive elements.

STANDARDS AUSTRALIA

Australian Standard Safety in laboratories

Part 4: Ionizing radiations

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE This Standard sets out the precautions needed to prevent unnecessary exposure of persons using sources of ionizing radiation in laboratories and other persons who could be harmed by accidental or planned releases of radioactive substances or external beams of radiation. It also describes the important characteristics of ionizing radiation-producing substances and apparatus, the nature of the hazards, laboratory design requirements, and other essential radiation protection information.

It is recommended that the requirements of this Standard be implemented in all laboratories in which—

(a) radioactive substances are used or stored; or

(b) irradiating apparatus is operated.

As a guide to good practice, persons operating with radioactive substances in less than licensable quantities are invited to follow the principles contained in this Standard.

NOTE: The exposure of members of the public is considered in Appendix A.

1.2 REFERENCED DOCUMENTS A list of referenced documents and other related publications is provided in Appendix B.

1.3 DEFINITIONS For the purpose of this Standard, the definitions below apply.

1.3.1 Absorbed dose—the energy absorbed by matter from ionizing radiation per unit mass of irradiated substance. The SI unit of absorbed dose is the joule per kilogram, with the special name gray (Gy). For radiation protection purposes, the absorbed dose is averaged over a tissue or organ.

1.3.2 Activity—the average number of spontaneous nuclear transformations of a radionuclide occurring in unit time. The SI unit of activity is the becquerel (Bq), which is equal to one nuclear transformation per second.

1.3.3 Activity median aerodynamic diameter (AMAD)—The aerodynamic diameter (see Clause 1.3.5) of particles having the median activity of the aerosol of interest. That is, 50% of the aerosol activity is associated with larger particles. For dosimetry purposes, the AMAD can be taken to represent the aerosol.

1.3.4 Adequate protection—protection against ionizing radiations so that the radiation doses received by any person from external or internal sources, or both, are as low as is reasonably achievable (the ALARA principle), are below any constraints recommended for those sources by the National Health and Medical Research Council (NHMRC) and, in any event, do not exceed the maxima referred to in Section 4.

1.3.5 Aerodynamic diameter—the diameter, in μ m, of a unit density sphere having the same aerodynamic characteristics as the particle of interest.



The remainder of this document is available for purchase online at <u>www.saiglobal.com/shop</u>

SAI Global also carries a wide range of publications from a wide variety of Standards Publishers:

















Click on the logos to search the database online.